

EU DECLARATION OF CONFORMITY

Kamstrup A/S
Industrivej 28, Stilling
DK-8660 Skanderborg
T: +45 89 93 10 00
F: +45 89 93 10 01
M: kamstrup@kamstrup.com

We, Kamstrup A/S declare under our sole responsibility that the product

Product name: ULTRAFLOW® 54 DN150...300 mm, q_p 150...1000 m³/h

Product description: A flow sensor (4-transducer) used as a subassembly for a thermal energy meter

Type number: 65-5-XXXX-XXX
65-5-XXXX-XX-XXX

Serial number: 2017/XXXXXXXX

Conforms to the European directives as marked below:

Electromagnetic Compatibility Directive (EMC)
2014/30/EU

Standards:

- EN 61000-6-2:2005 EMC Immunity for industrial environments
- EN 61000-6-3:2007 EMC Emission for residential, commercial & light industrial environments

Low Voltage Directive (LVD)
2014/35/EU

Standard:

- EN 61010-1:2010

Pressure Equipment Directive (PED)

2014/68/EU

DN sizes in the table applies:

Fluid	Group 2	
Category	II	
Module	A2	
Maximum allowable pressure PS	25	16
Nominal bore	DN150-DN250	DN300
Notified body, Inspection	Force Certification (0200) Park Allé 345 DK-2605 Brøndby	
Notified body, Quality system ISO 9001	DNV GL Tuborg Parkvej 8, 2. DK-2900 Hellerup	

Conformity assessment procedures according to directive 2014/68/EU, Annex III.

Measuring Instrument Directive (MID)

2014/32/EU

Applies for products marked with CE marking and supplementary metrology marking according to Article 21 in directive 2014/32/EU.

Standards:

- EN 1434-4:2007
- Welmec 7.2

Type certificate number: DK-0200-MI004-008

Notified Body, Module D Certificate:

Force Certification A/S
EU Notified Body number: 0200
Park Allé 345, 2605 Brøndby
Denmark

RoHS Directive (Restriction of Hazardous Substances)

2011/65/EU

References:

Kamstrup procedure 5509-016, linked with supplier statement in 5509-045.

Date: 2016-12-23, Skanderborg**Quality Assurance Manager:** _____


Viggo Andersen